

# Certificate of Participation

for the EURADOS Intercomparison 2022 for whole body dosimeters (IC2022ph)

**Certificate Number:** EURADOS-IC2022ph-S106 for system S106/2022  
**Number of pages:** 4  
**Date of Issue:** July 07<sup>th</sup>, 2023  
**Participating Institute:** Research and Production Enterprise DOSIMETRICA LLC, Ukraine  
**Dosimetry System:** Harshaw 8800, s/n 112053  
**Reporting number:** 18 (this anonymous number will be used in further publications)  
**Intercomparison procedure:** The EURADOS Intercomparison 2022 for whole body dosimeters was managed and coordinated on behalf of EURADOS by the WG2 Intercomparison Organization Group (OG). The OG established the irradiation plan and announced the intercomparison, including the range limits of the doses and radiation qualities, in May 2022.  
Participants were asked to indicate details of the dosimeter reference point on the online application form. After completing application procedures the participants sent their dosimeters, according to the instructions, to the OG coordinator (July 2022). The coordinator checked the correct labelling of the dosimeters and transferred all dosimeters, along with the instructions, to the irradiation laboratory. The laboratory irradiated the dosimeters according to the irradiation plan and sent all the dosimeters back to the coordinator (November 2022).  
The coordinator then returned the dosimeters to the participants for assessment and indicated which dosimeters were not irradiated. The participants were instructed to follow normal routine procedures as far as possible. The participants then sent the results of the dosimeter readings to the coordinator (January 2023). After receipt of the participants' results, the coordinator sent the irradiation data to the participants.

**Number of participants:** 96 institutes participated in IC2022ph with a total of 116 systems.  
**Coordinator:** Christian Gärtner (Seibersdorf Labor GmbH, A-2444 Seibersdorf)  
**Intercomparison results:** See the tables on pages 2 to 4 of this certificate.  
**Irradiation data:** See the attached certificate of the irradiation laboratory:  
Number Dos/2896-106/2022  
**Participant results:** See the attached signed dose report provided by the participant.

On behalf of the intercomparison  
Organization Group:



Christian Gärtner  
Coordinator

On behalf of EURADOS:



Filip Vanhavere  
Chairperson

## Whole body dosemeter intercomparison IC2022ph

### Result of the Intercomparison (Dosimetry System S106/2022)

EURADOS Dosemeter ID	Participant's Dosemeter ID	Radiation Quality	Quantity	Participant's Value	Reference Value	Ratio
S106/2022-03	400003	S-Cs, 0°	$H_p(10)$	2.77 mSv	2.60 mSv	1.07
			$H_p(0.07)$	2.79 mSv	2.60 mSv	1.07
S106/2022-10	400010	S-Cs, 0°	$H_p(10)$	2.77 mSv	2.60 mSv	1.07
			$H_p(0.07)$	2.89 mSv	2.60 mSv	1.11
S106/2022-12	400013	S-Cs, 0°	$H_p(10)$	2.74 mSv	2.60 mSv	1.05
			$H_p(0.07)$	2.74 mSv	2.60 mSv	1.05
S106/2022-15	400016	S-Cs, 0°	$H_p(10)$	2.69 mSv	2.60 mSv	1.03
			$H_p(0.07)$	2.70 mSv	2.60 mSv	1.04
S106/2022-04	400004	S-Cs, 0°	$H_p(10)$	8.03 mSv	7.50 mSv	1.07
			$H_p(0.07)$	8.47 mSv	7.50 mSv	1.13
S106/2022-07	400007	S-Cs, 0°	$H_p(10)$	7.98 mSv	7.50 mSv	1.06
			$H_p(0.07)$	7.94 mSv	7.50 mSv	1.06
S106/2022-14	400015	S-Co, 0°	$H_p(10)$	10.15 mSv	9.64 mSv	1.05
			$H_p(0.07)$	10.33 mSv	9.81 mSv	1.05
S106/2022-26	400029	S-Co, 0°	$H_p(10)$	10.08 mSv	9.64 mSv	1.05
			$H_p(0.07)$	11.00 mSv	9.81 mSv	1.12
S106/2022-05	400005	S-Co, 0°	$H_p(10)$	140.00 mSv	131.00 mSv	1.07
			$H_p(0.07)$	155.90 mSv	133.30 mSv	1.17
S106/2022-25	400027	S-Co, 0°	$H_p(10)$	139.06 mSv	131.00 mSv	1.06
			$H_p(0.07)$	156.24 mSv	133.30 mSv	1.17

## Whole body dosemeter intercomparison IC2022ph

### Result of the Intercomparison (Dosimetry System S106/2022), continued

EURADOS Dosemeter ID	Participant's Dosemeter ID	Radiation Quality	Quantity	Participant's Value	Reference Value	Ratio
S106/2022-06	400006	S-Co, 0°	$H_p(10)$	361.37 mSv	319.00 mSv	1.13
			$H_p(0.07)$	401.19 mSv	324.50 mSv	1.24
S106/2022-29	400032	S-Co, 0°	$H_p(10)$	366.33 mSv	319.00 mSv	1.15
			$H_p(0.07)$	380.35 mSv	324.50 mSv	1.17
S106/2022-22	400024	N-40, 0°	$H_p(10)$	6.06 mSv	5.80 mSv	1.04
			$H_p(0.07)$	6.20 mSv	6.14 mSv	1.01
S106/2022-23	400025	N-40, 0°	$H_p(10)$	6.08 mSv	5.80 mSv	1.05
			$H_p(0.07)$	6.18 mSv	6.14 mSv	1.01
S106/2022-18	400019	N-40 + S-Cs mixed, 0°	$H_p(10)$	3.42 mSv	3.00 mSv	1.14
			$H_p(0.07)$	3.52 mSv	3.10 mSv	1.14
S106/2022-21	400023	N-40 + S-Cs mixed, 0°	$H_p(10)$	3.50 mSv	3.00 mSv	1.17
			$H_p(0.07)$	3.50 mSv	3.10 mSv	1.13
S106/2022-01	400001	W-80, 0°	$H_p(10)$	5.88 mSv	6.00 mSv	0.98
			$H_p(0.07)$	6.01 mSv	5.56 mSv	1.08
S106/2022-13	400014	W-80, 0°	$H_p(10)$	5.72 mSv	6.00 mSv	0.95
			$H_p(0.07)$	6.23 mSv	5.56 mSv	1.12
S106/2022-02	400002	W-80, 60°	$H_p(10)$	5.93 mSv	5.80 mSv	1.02
			$H_p(0.07)$	5.80 mSv	6.26 mSv	0.93
S106/2022-17	400018	W-80, 60°	$H_p(10)$	6.61 mSv	5.80 mSv	1.14
			$H_p(0.07)$	6.87 mSv	6.26 mSv	1.10

## Whole body dosimeter intercomparison IC2022ph

### Result of the Intercomparison (Dosimetry System S106/2022), continued

EURADOS Dosemeter ID	Participant's Dosemeter ID	Radiation Quality
S106/2022-08	400008	not irradiated
S106/2022-09	400009	not irradiated
S106/2022-11	400011	not irradiated
S106/2022-16	400017	not irradiated
S106/2022-19	400020	not irradiated
S106/2022-20	400022	not irradiated
S106/2022-24	400026	not irradiated
S106/2022-27	400030	not irradiated
S106/2022-28	400031	not irradiated
S106/2022-30	400034	not irradiated

#### Radiation Qualities and average photon energy (according to ISO 4037-1 and IEC 61267):

- Nuclide Radiation:
  - S-Cs: 662 keV
  - S-Co: 1250 keV
- X-Rays:
  - N-40: 33 keV
  - W-80: 56.5 keV



Calibrations  
Cert. No 116(s)

HELLENIC REPUBLIC  
MINISTRY OF DEVELOPMENT AND INVESTMENTS  
GENERAL SECRETARIAT FOR RESEARCH & INNOVATION

Ag. Paraskevi, 19/05/2023  
Our Ref: B/428/11274



IONIZING RADIATION CALIBRATION LABORATORY  
Affiliated to the Hellenic Metrology Institute

IRRADIATION CERTIFICATE No: Dos/2896- 106/2022

Number of Pages: 2

Date of Issue: 24/05/2023

The following personnel dosimeters from: **EURADOS INTERCOMPARISON PROGRAM  
SYSTEM No:S106/2022**

have been irradiated at the *Ionizing Radiation Calibration Laboratory of Greek Atomic Energy Commission:*

<b>Personal Dosimeters (PD):</b>	<b>Whole body</b>
<b>Dosimeter Identification:</b>	-
<b>Detection Principle:</b>	-
<b>Irradiation Period:</b>	<b>See below</b>

The Kair reference values have been obtained using the secondary standards ionization chambers PTW W-32002-LS01 (S/N: 69) or FC65-G (S/N: 634) and the electrometer PTW UNIDOS 10002 (S/N 20314). The LS01 chamber was calibrated in PTB during 06 & 07-03-2019 (PTB, Cal. Cert. No PTB-6.3-4094018). The FC65-G chamber was calibrated at BIPM on 24-04-2019 (BIPM, Cal. Cert. No 25). The irradiation conditions are in accordance to ISO 4037/1-2-3-4 and IEC 62387.

**Irradiation conditions**

<b>Phantom:</b>	ISO water phantom, (30x30x15) cm <sup>3</sup>
<b>Source to PD Distance:</b>	100-300 cm, depending on required Kair rate
<b>Kair Rate:</b>	S-Cs: 454.3 µGy/min (at 100 cm) W-80: 2.63 mGy/min (at 200 cm) N-40: 203.7 µGy/min (at 200 cm) S-Co: 145.5 mGy/min (at 100 cm) S-Co: 18.6 mGy/min (at 300 cm) S-Co: 0.90 mGy/min (at 300 cm with lead block)
<b>Field Size:</b>	S-Cs: Circular with diameter of 55.6 cm (at 200 cm) x-rays: Circular with diameter 26.8 cm (at 200 cm) S-Co: Rectangular (30x30) cm <sup>2</sup> (at 300 cm)
<b>Build up PMMA:</b>	(0.3 x 30x30) cm <sup>3</sup>
<b>Reference point of PD:</b>	Frontal surface of slab phantom

**Environmental conditions during irradiations:**

Temperature	Pressure	Relative Humidity
19.0-21.0 °C	981.0-985.0 hPa	10 %

T.0. 60092, Agia Paraskevi 153 10 Attiki, Tel. : +30 210 650 6765  
e-mail : argiro.boziari@eeae.gr, info@eeae.gr

**Irradiation Data**

# Dosimeter	Date	Quality	H <sub>p</sub> (10) mSv	U % <sup>(1)</sup>	H <sub>p</sub> (0.07) mSv	U % <sup>(1)</sup>
S106-12	02/09/2022	S-Cs	2.60	4.9	2.60	4.9
S106-03	02/09/2022	S-Cs	2.60	4.9	2.60	4.9
S106-10	02/09/2022	S-Cs	2.60	4.9	2.60	4.9
S106-15	02/09/2022	S-Cs	2.60	4.9	2.60	4.9
S106-07	25/08/2022	S-Cs	7.50	4.9	7.50	4.9
S106-04	25/08/2022	S-Cs	7.50	4.9	7.50	4.9
S106-14	18/11/2022	S-Co	9.64	4.9	9.81	4.9
S106-26	18/11/2022	S-Co	9.64	4.9	9.81	4.9
S106-25	26/09/2022	S-Co	131	4.9	133	4.9
S106-05	26/09/2022	S-Co	131	4.9	133	4.9
S106-29	29/08/2022	S-Co	319	4.9	325	4.9
S106-06	29/08/2022	S-Co	319	4.9	325	4.9
S106-22	03/10/2022	N-40-(0°)	5.80	5.1	6.14	5.1
S106-23	03/10/2022	N-40-(0°)	5.80	5.1	6.14	5.1
S106-01	08/11/2022	W-80-(0°)	6.00	5.1	5.56	5.1
S106-13	08/11/2022	W-80-(0°)	6.00	5.1	5.56	5.1
S106-17	14/11/2022	W-80-(60°)	5.80	5.3	6.26	5.3
S106-02	14/11/2022	W-80-(60°)	5.80	5.3	6.26	5.3

# Dosimeter	Date	Quality	H <sub>p</sub> (10) mSv	U % <sup>(1)</sup>	H <sub>p</sub> (0.07) mSv	U % <sup>(1)</sup>
S106-18	12/09/2022	Cs-137	1.20	4.9 %	1.20	4.9 %
	31/10/2022	N-40	1.80	5.1 %	1.87	5.1 %
	Total		3.00	7.1 %	3.07	7.1 %
S106-21	12/09/2022	Cs-137	1.20	4.9 %	1.20	4.9 %
	31/10/2022	N-40	1.80	5.1 %	1.87	5.1 %
	Total		3.00	7.1 %	3.07	7.1 %

<sup>1</sup>U= uncertainty 95% confidence level (k=2)

**Not Irradiated Dosimeters :**

S106-08; S106-09; S106-11; S106-16; S106-19; S106-20; S106-24; S106-27; S106-28; S106-30;

Irradiations performed by:

Boziari A., Medical Physicist  
 Stamatopoulou E., Technician  
 Askounis P., Physicist



Argiro Boziari  
 Head of Dosimetry and Calibration Unit



This certificate is issued in accordance with the requirements of ISO 17025. It provides traceability of measurements to recognized national standards laboratories. The HIRCL/GAEC is a member of the IAEA/WHO Secondary Standard Dosimetry Laboratory Network. This certificate may not be reproduced other than in full, except with the prior written approval of the HIRCL/GAEC.

## Whole body dosimeter intercomparison IC2022ph Dose Values Form

*General Data*

**P028/2022, Research and Production Enterprise DOSIMETRICA LLC, 04050 ...**  
**S106/2022, Harshaw 8800, s/n 112053**

*Dose Values*

<b>Dosemeter 'S106/2022 - 01':</b>	<b>Hp(10) dose = 5.88 mSv</b>
400001	Hp(0.07) dose = 6.01 mSv
irradiated	Remark =
<b>Dosemeter 'S106/2022 - 02':</b>	<b>Hp(10) dose = 5.93 mSv</b>
400002	Hp(0.07) dose = 5.8 mSv
irradiated	Remark =
<b>Dosemeter 'S106/2022 - 03':</b>	<b>Hp(10) dose = 2.77 mSv</b>
400003	Hp(0.07) dose = 2.79 mSv
irradiated	Remark =
<b>Dosemeter 'S106/2022 - 04':</b>	<b>Hp(10) dose = 8.03 mSv</b>
400004	Hp(0.07) dose = 8.47 mSv
irradiated	Remark =
<b>Dosemeter 'S106/2022 - 05':</b>	<b>Hp(10) dose = 140 mSv</b>
400005	Hp(0.07) dose = 155.9 mSv
irradiated	Remark =
<b>Dosemeter 'S106/2022 - 06':</b>	<b>Hp(10) dose = 361.37 mSv</b>
400006	Hp(0.07) dose = 401.19 mSv
irradiated	Remark =
<b>Dosemeter 'S106/2022 - 07':</b>	<b>Hp(10) dose = 7.98 mSv</b>
400007	Hp(0.07) dose = 7.94 mSv
irradiated	Remark =
<b>Dosemeter 'S106/2022 - 08':</b>	<b>Hp(10) dose = 0.215 mSv</b>
400008	Hp(0.07) dose = 0.225 mSv
not irradiated	Remark =
<b>Dosemeter 'S106/2022 - 09':</b>	<b>Hp(10) dose = 0.224 mSv</b>
400009	Hp(0.07) dose = 0.225 mSv
not irradiated	Remark =
<b>Dosemeter 'S106/2022 - 10':</b>	<b>Hp(10) dose = 2.77 mSv</b>
400010	Hp(0.07) dose = 2.89 mSv
irradiated	Remark =



## Whole body dosimeter intercomparison IC2022ph Dose Values Form

**General Data**

P028/2022, Research and Production Enterprise DOSIMETRICA LLC, 04050 ...

S106/2022, Harshaw 8800, s/n 112053

**Dose Values, continued**

Dosemeter 'S106/2022 - 11': 400011 not irradiated	Hp(10) dose = 0.217 mSv Hp(0.07) dose = 0.216 mSv Remark =
Dosemeter 'S106/2022 - 12': 400013 irradiated	Hp(10) dose = 2.74 mSv Hp(0.07) dose = 2.74 mSv Remark =
Dosemeter 'S106/2022 - 13': 400014 irradiated	Hp(10) dose = 5.72 mSv Hp(0.07) dose = 6.23 mSv Remark =
Dosemeter 'S106/2022 - 14': 400015 irradiated	Hp(10) dose = 10.15 mSv Hp(0.07) dose = 10.33 mSv Remark =
Dosemeter 'S106/2022 - 15': 400016 irradiated	Hp(10) dose = 2.69 mSv Hp(0.07) dose = 2.7 mSv Remark =
Dosemeter 'S106/2022 - 16': 400017 not irradiated	Hp(10) dose = 0.221 mSv Hp(0.07) dose = 0.23 mSv Remark =
Dosemeter 'S106/2022 - 17': 400018 irradiated	Hp(10) dose = 6.61 mSv Hp(0.07) dose = 6.87 mSv Remark =
Dosemeter 'S106/2022 - 18': 400019 irradiated	Hp(10) dose = 3.42 mSv Hp(0.07) dose = 3.52 mSv Remark =
Dosemeter 'S106/2022 - 19': 400020 not irradiated	Hp(10) dose = 0.222 mSv Hp(0.07) dose = 0.22 mSv Remark =
Dosemeter 'S106/2022 - 20': 400022 not irradiated	Hp(10) dose = 0.219 mSv Hp(0.07) dose = 0.211 mSv Remark =



## Whole body dosimeter intercomparison IC2022ph Dose Values Form

*General Data*

**P028/2022, Research and Production Enterprise DOSIMETRICA LLC, 04050 ...**  
**S106/2022, Harshaw 8800, s/n 112053**

*Dose Values, continued*

<b>Dosemeter 'S106/2022 - 21':</b>	Hp(10) dose = 3.5 mSv
400023	Hp(0.07) dose = 3.5 mSv
irradiated	Remark =
<b>Dosemeter 'S106/2022 - 22':</b>	Hp(10) dose = 6.06 mSv
400024	Hp(0.07) dose = 6.2 mSv
irradiated	Remark =
<b>Dosemeter 'S106/2022 - 23':</b>	Hp(10) dose = 6.08 mSv
400025	Hp(0.07) dose = 6.18 mSv
irradiated	Remark =
<b>Dosemeter 'S106/2022 - 24':</b>	Hp(10) dose = 0.218 mSv
400026	Hp(0.07) dose = 0.23 mSv
not irradiated	Remark =
<b>Dosemeter 'S106/2022 - 25':</b>	Hp(10) dose = 139.06 mSv
400027	Hp(0.07) dose = 156.24 mSv
irradiated	Remark =
<b>Dosemeter 'S106/2022 - 26':</b>	Hp(10) dose = 10.08 mSv
400029	Hp(0.07) dose = 11 mSv
irradiated	Remark =
<b>Dosemeter 'S106/2022 - 27':</b>	Hp(10) dose = 0.218 mSv
400030	Hp(0.07) dose = 0.218 mSv
not irradiated	Remark =
<b>Dosemeter 'S106/2022 - 28':</b>	Hp(10) dose = 0.213 mSv
400031	Hp(0.07) dose = 0.209 mSv
not irradiated	Remark =
<b>Dosemeter 'S106/2022 - 29':</b>	Hp(10) dose = 366.33 mSv
400032	Hp(0.07) dose = 380.35 mSv
irradiated	Remark =
<b>Dosemeter 'S106/2022 - 30':</b>	Hp(10) dose = 0.218 mSv
400034	Hp(0.07) dose = 0.232 mSv
not irradiated	Remark =

## Whole body dosimeter intercomparison IC2022ph Dose Values Form

*General Data*

P028/2022, Research and Production Enterprise DOSIMETRICA LLC, 04050 ...


S106/2022, Harshaw 8800, s/n 112053

*Dose Values, continued*

Transit dose: Hp(10) dose = 0.22 mSv

Remark =

  
Signature

  
Name

  
Date